



WeBIOPATR 2021

The Eighth International WEBIOPATR
Workshop & Conference
Particulate Matter: Research and Management

Abstracts of Keynote Invited Lectures and Contributed Papers

Milena Jovašević-Stojanović,
Alena Bartoňová,
Miloš Davidović and Simon Smith, Eds

Vinča Institute of Nuclear Sciences
Vinča, Belgrade 2021

**ABSTRACTS OF KEYNOTE INVITED LECTURES AND
CONTRIBUTED PAPERS**

The Eighth WeBIOPATR Workshop & Conference
Particulate Matter: Research and Management

WeBIOPATR 2021

29th November to 1st December 2021

Vinča, Belgrade, Serbia

Editors

Milena Jovašević-Stojanović

Alena Bartoňová

Miloš Davidović

Simon Smith

Publisher

Vinča Institute of Nuclear Sciences

Prof. Dr Snežana Pajović, Director

P.O.Box 522

11001 Belgrade, Serbia

Printed by

Vinča Institute of Nuclear Sciences

Number of copies

150

ISBN 978-86-7306-164-1

© Vinča Institute of Nuclear Sciences

Vinča, Belgrade 2021.

www.vin.bg.ac.rs/

Organizers



Vinča Institute of Nuclear Sciences, University of Belgrade, National Institute of the Republic of Serbia, Serbia

Public Health Institute of Belgrade, Serbia

NILU Norwegian Institute for Air Research, Norway

The 8th WeBIOPATR Workshop and Conference,

Particulate Matter: Research and Management, WEBIOPATR2021

is supported by:



Virtual centre for
Distributed atmospheric
Sensing for reduction of pollution pressures

*EC H2020 Framework Program for Research and Innovation,
area “Spreading excellence and widening participation”,
VIDIS project (2020-2023) coordinated by Vinča Institute of Nuclear Sciences,
Grant agreement number 952433.*



Ministry of Education, Science and Technological Development of the Republic of Serbia

11.8 PORTABLE AIR QUALITY MONITOR BASED ON LOW-COST SENSORS

V. Tasić (1), A. Božilov (2), N. Mišić (2), I. Lazović (3), N. Mirkov(3), D. Topalović (3)

(1) Mining and Metallurgy Institute Bor, Bor, Serbia, (2) Faculty of Occupational Safety Niš, University of Niš, Niš, Serbia (3) Institute Vinca, University of Belgrade, Belgrade, Serbia

visa.tasic@irmbor.co.rs

Background and Aims: National systems of automatic air quality monitoring are based on automatic measuring instruments that are installed in stationary air quality monitoring stations. Substantial financial resources are required for the procurement, installation, calibration, and maintenance of such measuring devices. For this reason, existing national and local air quality monitoring networks are not able to provide a high temporal and spatial resolution of measurement results. In the last ten years, as a possible alternative to the conventional approach to air quality monitoring, real-time monitoring systems using cheap (low-cost) sensors and sensor platforms have begun to be applied (Jovasevic-Stojanovic et al, 2015). This paper describes the basic characteristics of a portable air quality measuring device PAQMAN 2020 based on low-cost sensors. Part of the results of comparative measurements of this device with the measurement results of the reference instruments is presented in the paper.

Methods: PAQMAN 2020 uses an Arduino Mega microcontroller as a control board. An SDS011 sensor module is used for measurements of PM₁₀ and PM_{2.5} mass concentrations from the range of 0-2000 µg /m³. A DHT22 sensor module is used for temperature (-10 to +40 °C) and relative humidity (20 to 90 % RH) measurements. NDIR S8 CO₂ module is used for the measurement of CO₂ concentrations in the range from 400 to 2000 ppm. The measurement results are stored in text files on the microSD card and displayed on the device's LCD display (LCD2004). Device programming was done in C through the Arduino IDE. Results of measurement could be downloaded from PAQMAN 2020 to PC over a standard USB serial port.

Key results: The PAQMAN 2020 was tested in the laboratory and in the field. Mean 15-minute or mean hourly values were compared, depending on the available time resolution of the results from the reference instruments. Figure 1 show comparison of PM concentrations (SDS011 as DUT vs ref.) measured in the laboratory. Determination coefficients ($R^2 = 0.957$ for PM₁₀ and $R^2 = 0.945$ PM_{2.5}) indicate a very strong linear relationship between the measurement results of the reference instrument and the SDS 011 sensor.




Figure 1. Comparison of results for the PM concentrations measured in the laboratory

Conclusions: The PAQMAN 2020 monitor was showed very good stability and reliability during the test period. So that, it could be applied for temporary or continuous indicative measurements of indoor and outdoor air quality.

Acknowledgements: This work was financially supported by the Ministry of Education, Science and Technological Development of the Republic of Serbia, Grant No. 451-03-9/2021-14/ 200052.

REFERENCES

Jovasevic-Stojanovic M., Bartonova A., Topalovic D., Lazovic I., Pokric B., Ristovski Z., 2015. On the use of small and cheaper sensors and devices for indicative citizen-based monitoring of respirable particulate matter, Environ. Pollut., 206, 696–704.

CIP - Каталогизација у публикацији
Народна библиотека Србије, Београд

502.3:502.175(082)(0.034.2)

613.15(082)(0.034.2)

66.071.9(082)(0.034.2)

INTERNATIONAL WeBIOPATR Workshop Particulate Matter: Research and Management (8 ; 2021 ; Vinča) Abstracts of keynote invited lectures and contributed papers [Elektronski izvor] / The Eighth International WEBIOPATR Workshop & Conference Particulate Matter: Research and Management, WeBIOPATR 2021, 29th November to 1st December 2021 Vinča, Belgrade, Serbia ; [organizers Vinča Institute of Nuclear Sciences, University of Belgrade, National Institute of the Republic of Serbia [and] Public Health Institute of Belgrade, Serbia [and] NILU Norwegian Institute for Air Research, Norway] ; Milena Jovašević-Stojanović ... [et al.], eds. - Belgrade : Vinča Institute of Nuclear Sciences, 2021 (Belgrade : Vinča Institute of Nuclear Sciences). - 1 elektronski optički disk (DVD) ; 12 cm

Systemski zahtevi: Nisu navedeni. - Nasl. sa naslovne strane dokumenta. - "... Conference ... as a combination of online and face-to-face event." --> Preface. - Tiraž 150. - Preface / Milena Jovašević-Stojanović and Alena Bartoňová. - Bibliografija uz većinu apstrakata. - Registar.

ISBN 978-86-7306-164-1

1. International Conference Particulate Matter: Research and Management (8 ; 2021 ; Vinča)

а) Ваздух -- Контрола квалитета -- Зборници

б) Здравље -- Заштита -- Зборници

в) Отпадни гасови -- Штетно дејство -- Зборници

COBISS.SR-ID 53342985

ISBN 978-86-7306-164-1

